The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/1-H
S/N: 61145

Normalized Magnitude-frequency Characteristics

Freq. (Hz)	X-Axis (Horizontal)
1.00	1.0018
10.00	1.0000
20.00	0.9950
30.00	0.9878
40.00	0.9796
50.00	0.9725
60.00	0.9674
70.00	0.9664
80.00	0.9707
90.00	0.9811
100.00	0.9985
120.00	1.0562
140.00	1.1267
160.00	1.1269

Output Sensitivity

	X-Axis (Horizontal)
Output Sensitivity (V/g)	2.501

* g = 9.80 m/s^2