The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/1-V
S/N: 62051

Normalized Magnitude-frequency Characteristics

Freq. (Hz)	Z-Axis (Vertical)
1.00	1.0015
10.00	1.0000
20.00	0.9957
30.00	0.9895
40.00	0.9825
50.00	0.9763
60.00	0.9719
70.00	0.9707
80.00	0.9739
90.00	0.9820
100.00	0.9957
120.00	1.0404
140.00	1.0952
160.00	1.1046

Output Sensitivity

	Z-Axis (Vertical)
Output Sensitivity (V/g)	2.499

* g = 9.80 m/s^2