The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/1-V
S/N: 62062

Normalized Magnitude-frequency Characteristics

Freq. (Hz)	Z-Axis (Vertical)
1.00	1.0017
10.00	1.0000
20.00	0.9952
30.00	0.9882
40.00	0.9801
50.00	0.9726
60.00	0.9669
70.00	0.9645
80.00	0.9666
90.00	0.9740
100.00	0.9875
120.00	1.0346
140.00	1.0964
160.00	1.1088

Output Sensitivity

	Z-Axis (Vertical)
Output Sensitivity (V/g)	2.501

* g = 9.80 m/s^2