



**The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.**

**Reftek Systems Inc. purchased the REF TEK product line in May 2020.**  
**Reftek Systems will honour existing product warranties and extended warranties.**

**ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:**

**Sales & Administrative:** [sales@reftek.com](mailto:sales@reftek.com)

**Support & Engineering:** [support@reftek.com](mailto:support@reftek.com)

**Accounting:** [accounting@reftek.com](mailto:accounting@reftek.com)

**Address:** 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

**Phone:** +1-902-444-7650

**Fax:** +1-902-444-7651

## REF TEK 131A-02/3 Test Data Sheet

Customer: C H E S A Job File 15264 Unit S/N: 0401

Unit Part #: 131A-02/3

Sensor Part #: SF1500SN.A

Axis	Scale Factor – (Volts/g) (+/-10V F.S.)	Sensor Component Serial #
Ch. 1	2.548	A010-007
Ch. 2	2.480	A007-285
Ch. 3	2.517	A007-275