

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3/INT Test Data Sheet

Customer: USGS **Job File 12751**

Sensor Part #:	**147-01/3/INT**

Sensor S/N #:	**10010**

DAS S/N: AD8A

Normalized Sensitivity

Freq.(Hz)	**Ch. 1 (Vert)**	**Ch. 2 (N-S)**	**Ch. 3 (E-W)**
1.00	1.0066	1.0072	1.0074
10.00	1.0000	1.0000	1.0000
20.00	0.9852	0.9835	0.9831
40.00	0.9527	0.9450	0.9445
60.00	0.9438	0.9288	0.9296
80.00	0.9585	0.9395	0.9415
99.00	0.9584	0.9475	0.9468
120.00	0.8634	0.8839	0.8675
140.00	0.6731	0.7148	0.6835
160.00	0.4787	0.5166	0.4860
Sensitivity (@10Hz)	**2.501v/g**	**2.496v/g**	**2.508v/g**