

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3/INT Test Data Sheet

Customer: USGS **Job File 12751**

Sensor Part #:	**147-01/3/INT**

Sensor S/N #:	**10013**

DAS S/N: AD34

Normalized Sensitivity

Freq.(Hz)	**Ch. 1 (Vert)**	**Ch. 2 (N-S)**	**Ch. 3 (E-W)**
1.00	1.0072	1.0067	1.0079
10.00	1.0000	1.0000	1.0000
20.00	0.9838	0.9851	0.9819
40.00	0.9475	0.9539	0.9388
60.00	0.9338	0.9505	0.9161
80.00	0.9433	0.9762	0.9146
99.00	0.9419	0.9862	0.9026
120.00	0.8491	0.8778	0.8081
140.00	0.6602	0.6602	0.6313
160.00	0.4682	0.4586	0.4526
Sensitivity (@10Hz)	**2.502v/g**	**2.498v/g**	**2.496v/g**