

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3/INT Test Data Sheet

Customer: USGS **Job File 13008**

Sensor Part #:	**147-01/3/INT**

Sensor S/N #:	**10029**

DAS S/N: AF2A

Normalized Sensitivity

Freq.(Hz)	**Ch. 1 (Vert)**	**Ch. 2 (N-S)**	**Ch. 3 (E-W)**
1.00	1.0072	1.0075	1.0079
10.00	1.0000	1.0000	1.0000
20.00	0.9843	0.9837	0.9829
40.00	0.9466	0.9455	0.9420
60.00	0.9273	0.9266	0.9202
80.00	0.9318	0.9308	0.9212
99.00	0.9325	0.9296	0.9224
120.00	0.8972	0.8576	0.8631
140.00	0.7450	0.7001	0.7155
160.00	0.5619	0.5187	0.5314
Sensitivity (@10Hz)	2.499 v/g	2.501 v/g	2.500 v/g