

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3/INT Test Data Sheet

Customer: USGS **Job File 13008**

Sensor Part #:	**147-01/3/INT**

Sensor S/N #:	**10030**

DAS S/N: AF25

Normalized Sensitivity

Freq.(Hz)	**Ch. 1 (Vert)**	**Ch. 2 (N-S)**	**Ch. 3 (E-W)**
1.00	1.0071	1.0068	1.0073
10.00	1.0000	1.0000	1.0000
20.00	0.9838	0.9844	0.9836
40.00	0.9450	0.9456	0.9437
60.00	0.9296	0.9265	0.9238
80.00	0.9389	0.9288	0.9242
99.00	0.9453	0.9316	0.9201
120.00	0.8690	0.8758	0.8534
140.00	0.7009	0.7349	0.6994
160.00	0.5012	0.5500	0.5192
Sensitivity (@10Hz)	**2.501 v/g**	**2.499 v/g**	**2.500 v/g**