

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3/INT Test Data Sheet

Customer: INDYNE **Job File 12958D**

Sensor Part #:	**147-01/3/INT**

Sensor S/N #:	**10032**

DAS S/N: AF23

Normalized Sensitivity

Freq.(Hz)	**Ch. 1 (Vert)**	**Ch. 2 (N-S)**	**Ch. 3 (E-W)**
1.00	1.0076	1.0064	1.0070
10.00	1.0000	1.0000	1.0000
20.00	0.9831	0.9864	0.9845
40.00	0.9427	0.9573	0.9501
60.00	0.9215	0.9521	0.9372
80.00	0.9146	0.9697	0.9472
99.00	0.9118	0.9781	0.9553
120.00	0.8349	0.8883	0.8847
140.00	0.6890	0.6983	0.7180
160.00	0.5153	0.4966	0.5197
Sensitivity (@10Hz)	**2.499 v/g**	**2.500 v/g**	**2.500 v/g**