

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3/INT Test Data Sheet

Customer: PT.Mitra Intimarga Job File 14134

Sensor Part #:	**147-01/3/INT**

Sensor S/N #:	**10051AS**

DAS S/N: B470

Normalized Sensitivity

Freq.(Hz)	Ch. 1 (Vert)	Ch. 2 (N-S)	Ch. 3 (E-W)
1.00	1.0063	1.0069	1.0059
10.00	1.0000	1.0000	1.0000
20.00	0.9857	0.9844	0.9867
40.00	0.9532	0.9493	0.9574
60.00	0.9434	0.9386	0.9519
80.00	0.9620	0.9579	0.9769
99.00	0.9795	0.9730	1.0026
120.00	0.9192	0.8965	0.9485
140.00	0.7452	0.7062	0.7642
160.00	0.5411	0.5027	0.5495
Sensitivity (@10Hz)	**2.498 v/g**	**2.501 v/g**	**2.499 v/g**

Zero

z1=-4.70449e+002 +0.00000e+000i

Poles

p1=-2.77023e+004 +0.00000e+000i

p2=-1.13268e+003 +0.00000e+000i

p3=-2.96983e+002 +7.75369e+002i

p4=-2.96983e+002 -7.75369e+002i

p5=-3.41272e+002 +0.00000e+000i