

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3 Test Data Sheet

Customer: USGS **Job File 12734**

Sensor Part #:	**147-01/3**

Sensor S/N: G10703AS

Normalized Sensitivity

Freq.(Hz)	Ch. 1 (Vert)	Ch. 2 (N-S)	Ch. 3 (E-W)
1.00	1.0070	1.0068	1.0068
10.00	1.0000	1.0000	1.0000
20.00	0.9845	0.9851	0.9849
40.00	0.9492	0.9508	0.9490
60.00	0.9376	0.9403	0.9348
80.00	0.9520	0.9589	0.9481
99.00	0.9586	0.9746	0.9623
120.00	0.8831	0.9114	0.9138
140.00	0.7021	0.7300	0.7546
160.00	0.5058	0.5244	0.5552
Sensitivity (@10Hz)	**2.502 v/g**	**2.501 v/g**	**2.499 v/g**