

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3/EXT Test Data Sheet

Customer: USGS **Job File 12750**

Unit Part #:	**147-01/3-EXT**	

Unit S/N: G10704AS

Normalized Sensitivity

Freq.(Hz)	**Ch. 1 (Vert)**	**Ch. 2 (N-S)**	**Ch. 3 (E-W)**
1.00	1.0062	1.0068	1.0064
10.00	1.0000	1.0000	1.0000
20.00	0.9843	0.9825	0.9835
40.00	0.9510	0.9436	0.9468
60.00	0.9419	0.9285	0.9323
80.00	0.9583	0.9433	0.9435
99.00	0.9651	0.9621	0.9518
120.00	0.8850	0.9124	0.8882
140.00	0.7011	0.7399	0.7203
160.00	0.5046	0.5309	0.5251
Sensitivity (@20Hz)	**2.500 v/g**	**2.499 v/g**	**2.503 v/g**