

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147/01-3 Test Data Sheet

Customer: USGS **Job File 12750**

Sensor Part #:	**147-01/3**

Sensor S/N: G10708AS

Normalized Sensitivity

Freq.(Hz)	**Ch. 1 (Vert)**	**Ch. 2 (N-S)**	**Ch. 3 (E-W)**
1.00	1.0066	1.0065	1.0070
10.00	1.0000	1.0000	1.0000
20.00	0.9841	0.9849	0.9825
40.00	0.9471	0.9504	0.9410
60.00	0.9333	0.9403	0.9223
80.00	0.9445	0.9576	0.9300
99.00	0.9492	0.9712	0.9352
120.00	0.8757	0.9052	0.8672
140.00	0.7018	0.7256	0.6974
160.00	0.5091	0.5226	0.5040
Sensitivity (@10 Hz)	**2.499 v/g**	**2.500 v/g**	**2.497 v/g**