

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3 Test Data Sheet

Customer: Chesapeake **Job File 13961**

Sensor Part #:	**147-01/3**

Sensor S/N: G10720AS

Normalized Sensitivity

Freq.(Hz)	Ch. 1 (Vert)	Ch. 2 (N-S)	Ch. 3 (E-W)
1.00	1.0071	1.0069	1.0080
10.00	1.0000	1.0000	1.0000
20.00	0.9838	0.9848	0.9828
40.00	0.9473	0.9511	0.9448
60.00	0.9347	0.9411	0.9322
80.00	0.9501	0.9573	0.9443
99.00	0.9602	0.9622	0.9381
120.00	0.8846	0.8754	0.8282
140.00	0.6979	0.6846	0.6307
160.00	0.4980	0.4889	0.4438
Sensitivity (@10Hz)	**2.500 v/g**	**2.499 v/g**	**2.498 v/g**

Zero

z1 = -2.82101e +002

Poles

p1 = -3.31203e +002 +7.83946e +002i

p2 = -3.31203e +002 -7.83946e +002i

p3 = -7.56430e +002

p4 = -2.48145e +002