

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147/01-3 Test Data Sheet

Customer: USGS **Job File 12750**

Sensor Part #:	**147-01/3**

Sensor S/N: G10727AS

Normalized Sensitivity

Freq.(Hz)	**Ch. 1 (Vert)**	**Ch. 2 (N-S)**	**Ch. 3 (E-W)**
1.00	1.0069	1.0075	1.0066
10.00	1.0000	1.0000	1.0000
20.00	0.9847	0.9829	0.9850
40.00	0.9499	0.9452	0.9520
60.00	0.9401	0.9356	0.9468
80.00	0.9559	0.9524	0.9707
99.00	0.9616	0.9484	0.9821
120.00	0.8791	0.8325	0.8898
140.00	0.6925	0.6244	0.6835
160.00	0.4955	0.4348	0.4787
Sensitivity (@10 Hz)	**2.501 v/g**	**2.500 v/g**	**2.500 v/g**