

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3 Test Data Sheet

Customer: USGS **Job File 12734**

Sensor Part #:	**147-01/3**

Sensor S/N: G10730AS

Normalized Sensitivity

Freq.(Hz)	**Ch. 1 (Vert)**	**Ch. 2 (N-S)**	**Ch. 3 (E-W)**
1.00	1.0067	1.0073	1.0069
10.00	1.0000	1.0000	1.0000
20.00	0.9844	0.9826	0.9837
40.00	0.9502	0.9439	0.9480
60.00	0.9391	0.9278	0.9361
80.00	0.9543	0.9374	0.9512
99.00	0.9637	0.9426	0.9599
120.00	0.8888	0.8706	0.8803
140.00	0.7072	0.6968	0.6941
160.00	0.5094	0.5037	0.4956
Sensitivity (@10Hz)	**2.501 v/g**	**2.500 v/g**	**2.501 v/g**