

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147/01-3 Test Data Sheet

Customer: USGS **Job File 12750**

Sensor Part #:	**147-01/3**

Sensor S/N: G10734AS

Normalized Sensitivity

Freq.(Hz)	**Ch. 1 (Vert)**	**Ch. 2 (N-S)**	**Ch. 3 (E-W)**
1.00	1.0076	1.0070	1.0076
10.00	1.0000	1.0000	1.0000
20.00	0.9827	0.9843	0.9828
40.00	0.9422	0.9488	0.9436
60.00	0.9238	0.9372	0.9297
80.00	0.9321	0.9531	0.9466
99.00	0.9396	0.9643	0.9636
120.00	0.8764	0.8903	0.8974
140.00	0.7079	0.7037	0.7094
160.00	0.5125	0.5018	0.5014
Sensitivity (@10 Hz)	**2.503 v/g**	**2.499 v/g**	**2.500 v/g**