

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147/01-3 Test Data Sheet

Customer: USGS **Job File 12750**

Sensor Part #:	**147-01/3**

Sensor S/N: G10736AS

Normalized Sensitivity

Freq.(Hz)	**Ch. 1 (Vert)**	**Ch. 2 (N-S)**	**Ch. 3 (E-W)**
1.00	1.0072	1.0069	1.0073
10.00	1.0000	1.0000	1.0000
20.00	0.9831	0.9840	0.9824
40.00	0.9432	0.9478	0.9416
60.00	0.9276	0.9380	0.9261
80.00	0.9368	0.9549	0.9346
99.00	0.9381	0.9617	0.9312
120.00	0.8560	0.8724	0.8381
140.00	0.6746	0.6756	0.6523
160.00	0.4833	0.4772	0.4650
Sensitivity (@10 Hz)	**2.500 v/g**	**2.500 v/g**	**2.499 v/g**