

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147/01-3 Test Data Sheet

Customer: USGS **Job File 12750**

Sensor Part #:	**147-01/3**

Sensor S/N: G10737AS

Normalized Sensitivity

Freq.(Hz)	**Ch. 1 (Vert)**	**Ch. 2 (N-S)**	**Ch. 3 (E-W)**
1.00	1.0066	1.0068	1.0059
10.00	1.0000	1.0000	1.0000
20.00	0.9827	0.9822	0.9848
40.00	0.9442	0.9431	0.9536
60.00	0.9286	0.9284	0.9482
80.00	0.9412	0.9412	0.9684
99.00	0.9545	0.9487	0.9730
120.00	0.8960	0.8704	0.8736
140.00	0.7361	0.7451	0.7230
160.00	0.5229	0.4871	0.4778
Sensitivity (@10 Hz)	**2.499 v/g**	**2.502 v/g**	**2.500 v/g**