

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3 Test Data Sheet

Customer: USGS **Job File 12734**

Sensor Part #:	**147-01/3**

Sensor S/N: G10743AS

Normalized Sensitivity

Freq.(Hz)	**Ch. 1 (Vert)**	**Ch. 2 (N-S)**	**Ch. 3 (E-W)**
1.00	1.0072	1.0067	1.0073
10.00	1.0000	1.0000	1.0000
20.00	0.9839	0.9852	0.9830
40.00	0.9469	0.9529	0.9421
60.00	0.9355	0.9134	0.9181
80.00	0.9338	0.9677	0.9239
99.00	0.9618	0.9783	0.9257
120.00	0.8896	0.8877	0.8645
140.00	0.7047	0.6870	0.7109
160.00	0.5019	0.4844	0.5249
Sensitivity (@10Hz)	**2.500 v/g**	**2.500 v/g**	**2.500 v/g**