

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3 Test Data Sheet

Customer: USGS **Job File 12733**

Sensor Part #:	**147-01/3**

Sensor S/N: G10751AS

Normalized Sensitivity

Freq.(Hz)	Ch. 1 (Vert)	Ch. 2 (N-S)	Ch. 3 (E-W)
1.00	1.0067	1.0060	1.0064
10.00	1.0000	1.0000	1.0000
20.00	0.9842	0.9858	0.9842
40.00	0.9512	0.9578	0.9496
60.00	0.9456	0.9575	0.9375
80.00	0.9724	0.9866	0.9532
99.00	0.9962	1.0021	0.9704
120.00	0.9206	0.9055	0.9167
140.00	0.7147	0.6925	0.7489
160.00	0.4994	0.4845	0.5446
Sensitivity (@10Hz)	**2.503 v/g**	**2.499 v/g**	**2.501 v/g**