

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3 Test Data Sheet

Customer: USGS **Job File 12734**

Sensor Part #:	**147-01/3**

Sensor S/N: G10786AS

Normalized Sensitivity

Freq.(Hz)	Ch. 1 (Vert)	Ch. 2 (N-S)	Ch. 3 (E-W)
1.00	1.0059	1.0059	1.0060
10.00	1.0000	1.0000	1.0000
20.00	0.9850	0.9855	0.9849
40.00	0.9566	0.9586	0.9546
60.00	0.9584	0.9608	0.9508
80.00	0.9901	0.9899	0.9705
99.00	1.0007	0.9935	0.9650
120.00	0.8823	0.8684	0.8434
140.00	0.6565	0.6466	0.6355
160.00	0.4532	0.4496	0.4461
Sensitivity (@10Hz)	**2.500 v/g**	**2.501 v/g**	**2.500 v/g**