

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3 Test Data Sheet

Customer: USGS **Job File 12734**

Sensor Part #:	**147-01/3**

Sensor S/N: G10789AS

Normalized Sensitivity

Freq.(Hz)	Ch. 1 (Vert)	Ch. 2 (N-S)	Ch. 3 (E-W)
1.00	1.0063	1.0067	1.0061
10.00	1.0000	1.0000	1.0000
20.00	0.9852	0.9843	0.9862
40.00	0.9517	0.9477	0.9551
60.00	0.9434	0.9360	0.9494
80.00	0.9606	0.9491	0.9689
99.00	0.9710	0.9565	0.9797
120.00	0.8938	0.8810	0.8969
140.00	0.7084	0.7009	0.7052
160.00	0.5090	0.5047	0.5040
Sensitivity (@10Hz)	**2.500 v/g**	**2.501 v/g**	**2.501 v/g**