

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3 Test Data Sheet

Customer: USGS **Job File 12734**

Sensor Part #:	**147-01/3**

Sensor S/N: G10803AS

Normalized Sensitivity

Freq.(Hz)	**Ch. 1 (Vert)**	**Ch. 2 (N-S)**	**Ch. 3 (E-W)**
1.00	1.0066	1.0065	1.0062
10.00	1.0000	1.0000	1.0000
20.00	0.9853	0.9856	0.9862
40.00	0.9534	0.9530	0.9567
60.00	0.9478	0.9427	0.9519
80.00	0.9723	0.9598	0.9730
99.00	0.9878	0.9772	0.9814
120.00	0.8956	0.9215	0.8884
140.00	0.6864	0.7512	0.6910
160.00	0.4795	0.5471	0.4921
Sensitivity (@10Hz)	**2.501 v/g**	**2.499 v/g**	**2.501 v/g**