

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3/EXT Test Data Sheet

Customer: Beijing **Job File 13226**

Sensor Part #:	**147-01/3**

Sensor S/N: G10807AS

Normalized Sensitivity

Freq.(Hz)	Ch. 1 (Vert)	Ch. 2 (N-S)	Ch. 3 (E-W)
1.00	1.0065	1.0060	1.0069
10.00	1.0000	1.0000	1.0000
20.00	0.9843	0.9846	0.9833
40.00	0.9487	0.9518	0.9456
60.00	0.9325	0.9389	0.9288
80.00	0.9389	0.9487	0.9372
99.00	0.9438	0.9538	0.9443
120.00	0.8882	0.8918	0.8845
140.00	0.7382	0.7312	0.7222
160.00	0.5524	0.5411	0.5272
Sensitivity (@10Hz)	**2.500 v/g**	**2.501 v/g**	**2.500 v/g**