


**The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.**

**Reftek Systems Inc. purchased the REF TEK product line in May 2020.**

**Reftek Systems will honour existing product warranties and extended warranties.**

**ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:**

**Sales & Administrative:** sales@reftek.com

**Support & Engineering:** support@reftek.com

**Accounting:** accounting@reftek.com

**Address:** 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

**Phone:** +1-902-444-7650

**Fax:** +1-902-444-7651

# REF TEK 147-01/3 Test Data Sheet

**Customer:** **Beijing** **Job File** **13226**

| **Sensor Part #:** | **147-01/3** |
|---|---|

**Sensor S/N: G10810AS**

## Normalized Sensitivity

| **Freq.(Hz)** | **Ch. 1 (Vert)** | **Ch. 2 (N-S)** | **Ch. 3 (E-W)** |
|---|---|---|---|
| 1.00 | 1.0062 | 1.0067 | 1.0060 |
| 10.00 | 1.0000 | 1.0000 | 1.0000 |
| 20.00 | 0.9856 | 0.9839 | 0.9861 |
| 40.00 | 0.9527 | 0.9460 | 0.9547 |
| 60.00 | 0.9409 | 0.9290 | 0.9462 |
| 80.00 | 0.9535 | 0.9384 | 0.9634 |
| 99.00 | 0.9611 | 0.9464 | 0.9714 |
| 120.00 | 0.9008 | 0.8899 | 0.8965 |
| 140.00 | 0.7383 | 0.7286 | 0.7159 |
| 160.00 | 0.5415 | 0.5361 | 0.5172 |
| **Sensitivity (@10Hz)** | **2.500 v/g** | **2.500 v/g** | **2.501 v/g** |