

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3 Test Data Sheet

Customer: **REF TEK Sales** **Job File** **13453**

Sensor Part #:	**147-01/3**

Sensor S/N: G10817AS

Normalized Sensitivity

Freq.(Hz)	Ch. 1 (Vert)	Ch. 2 (N-S)	Ch. 3 (E-W)
1.00	1.0065	1.0063	1.0063
10.00	1.0000	1.0000	1.0000
20.00	0.9859	0.9871	0.9867
40.00	0.9552	0.9585	0.9585
60.00	0.9495	0.9539	0.9546
80.00	0.9736	0.9758	0.9755
99.00	0.9921	0.9882	0.9793
120.00	0.9173	0.9108	0.8810
140.00	0.7213	0.7221	0.6839
160.00	0.5102	0.5203	0.4870
Sensitivity (@10Hz)	**2.500 v/g**	**2.500 v/g**	**2.501 v/g**

Zero

z1 = -2.82101e +002

Poles

p1 = -3.31203e +002 +7.83946e +002i

p2 = -3.31203e +002 -7.83946e +002i

p3 = -7.56430e +002

p4 = -2.48145e +002