

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3 Test Data Sheet

Customer: **CONCO** **Job File** **15288**

Sensor Part #:	**147-01/3**

Sensor S/N: G11704AS

Normalized Sensitivity

Freq.(Hz)	**Ch. 1 (Vert)**	**Ch. 2 (N-S)**	**Ch. 3 (E-W)**
1.00	1.0065	1.0074	1.0065
10.00	1.0000	1.0000	1.0000
20.00	0.9868	0.9843	0.9870
40.00	0.9586	0.9518	0.9610
60.00	0.9554	0.9478	0.9652
80.00	0.9808	0.9741	1.0012
99.00	1.0023	0.9911	1.0231
120.00	0.9373	0.8997	0.9155
140.00	0.7476	0.6787	0.6840
160.00	0.5336	0.4492	0.4694
Sensitivity (@10 Hz)	**2.500v/g**	**2.500v/g**	**2.500v/g**

Zero **Poles**

z1=-4.70449e+002 +0.00000e+000i p1=-2.77023e+004 +0.00000e+000i

p2=-1.13268e+003 +0.00000e+000i

p3=-2.96983e+002 +7.75369e+002i

p4=-2.96983e+002 -7.75369e+002i

p5=-3.41272e+002 +0.00000e+000i