

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3 Test Data Sheet

Customer: CONCO **Job File 15653**

Sensor Part #:	**147-01/3**

Sensor S/N: G11717AS

Normalized Sensitivity

Freq.(Hz)	Ch. 1 (Vert)	Ch. 2 (N-S)	Ch. 3 (E-W)
1.00	1.0060	1.0061	1.0060
10.00	1.0000	1.0000	1.0000
20.00	0.9847	0.9845	0.9851
30.00	0.9669	0.9662	0.9682
40.00	0.9526	0.9509	0.9553
50.00	0.9455	0.9428	0.9497
60.00	0.9467	0.9430	0.9522
70.00	0.9552	0.9513	0.9616
80.00	0.9693	0.9671	0.9763
90.00	0.9819	0.9844	0.9894
100.00	0.9847	0.9968	0.9931
120.00	0.9132	0.9560	0.9256
140.00	0.7179	0.7728	0.7328
160.00	0.5178	0.5847	0.5345
Sensitivity (@10 Hz)	**2.500 v/g**	**2.500 v/g**	**2.500 v/g**

Zero

z1=-4.70449e+002 +0.00000e+000i

Poles

p1=-2.77023e+004 +0.00000e+000i

p2=-1.13268e+003 +0.00000e+000i

p3=-2.96983e+002 +7.75369e+002i

p4=-2.96983e+002 -7.75369e+002i

p5=-3.41272e+002 +0.00000e+000i