

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

REF TEK 147-01/3 Test Data Sheet

Customer: CONCO **Job File 15653**

Sensor Part #:	**147-01/3**

Sensor S/N: G11720AS

Normalized Sensitivity

Freq.(Hz)	Ch. 1 (Vert)	Ch. 2 (N-S)	Ch. 3 (E-W)
1.00	1.0050	1.0053	1.0054
10.00	1.0000	1.0000	1.0000
20.00	0.9873	0.9869	0.9864
30.00	0.9727	0.9725	0.9709
40.00	0.9616	0.9622	0.9591
50.00	0.9578	0.9596	0.9546
60.00	0.9628	0.9656	0.9584
70.00	0.9753	0.9788	0.9695
80.00	0.9935	0.9972	0.9866
90.00	1.0083	1.0123	1.0022
100.00	1.0094	1.0145	1.0081
120.00	0.9157	0.9278	0.9390
140.00	0.6960	0.7132	0.7367
160.00	0.5022	0.5109	0.5315
Sensitivity (@10 Hz)	**2.500 v/g**	**2.500 v/g**	**2.500 v/g**

Zero

z1=-4.70449e+002 +0.00000e+000i

Poles

p1=-2.77023e+004 +0.00000e+000i

p2=-1.13268e+003 +0.00000e+000i

p3=-2.96983e+002 +7.75369e+002i

p4=-2.96983e+002 -7.75369e+002i

p5=-3.41272e+002 +0.00000e+000i