

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Tel : 214-440-1265
Fax : 972-578-0045
www.reftek.com

REF TEK
A DIVISION OF TRIMBLE

REF TEK 147-01/3 Test Data Sheet

SENSOR PART #	147-01/3
SENSOR S/N #	G11726AS

Normalized Sensitivity

Freq.(Hz)	Ch. 1 (Vert)	Ch. 2 (N-S)	Ch. 3 (E-W)
1.00	1.0048	1.0056	1.0053
10.00	1.0000	1.0000	1.0000
20.00	0.9879	0.9860	0.9865
30.00	0.9740	0.9697	0.9707
40.00	0.9638	0.9571	0.9578
50.00	0.9611	0.9518	0.9515
60.00	0.9675	0.9551	0.9527
70.00	0.9815	0.9659	0.9608
80.00	1.0012	0.9824	0.9742
90.00	1.0167	0.9965	0.9866
100.00	1.0170	0.9992	0.9910
120.00	0.9171	0.9191	0.9327
140.00	0.6943	0.7128	0.7546
160.00	0.4921	0.5212	0.5520
Sensitivity	2.500v/g	2.500v/g	2.500v/g

Zero

z1=-4.70449e+002 +0.00000e+000i

Poles

p1=-2.77023e+004 +0.00000e+000i
p2=-1.13268e+003 +0.00000e+000i
p3=-2.96983e+002 +7.75369e+002i
p4=-2.96983e+002 -7.75369e+002i
p5=-3.41272e+002 +0.00000e+000i