



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
 1600 Tenth Street
 Suite A
 Plano, Texas 75074
 Support: 1-888-879-2207
 Fax: 972-578-0045
www.trimble.com/infrastructure



REF TEK 147-01/3 Test Data Sheet

Sensor Part #:	147-01/3
-----------------------	-----------------

Sensor S/N #:	G11755AS
----------------------	-----------------

Normalized Sensitivity

Freq.(Hz)	Ch. 1 (Vert)	Ch. 2 (N-S)	Ch. 3 (E-W)
1.00	1.0056	1.0058	1.0055
10.00	1.0000	1.0000	1.0000
20.00	0.9858	0.9851	0.9859
30.00	0.9692	0.9675	0.9690
40.00	0.9557	0.9528	0.9547
50.00	0.9491	0.9454	0.9470
60.00	0.9500	0.9464	0.9471
70.00	0.9577	0.9556	0.9543
80.00	0.9703	0.9723	0.9672
90.00	0.9814	0.9898	0.9789
100.00	0.9835	1.0011	0.9820
120.00	0.9187	0.9533	0.9166
140.00	0.7372	0.7658	0.7308
160.00	0.5359	0.5712	0.5301
Sensitivity (@10 Hz)	2.500 v/g	2.500 v/g	2.500 v/g

Zero

z1=-4.70449e+002 +0.00000e+000i

Poles

p1=-2.77023e+004 +0.00000e+000i

p2=-1.13268e+003 +0.00000e+000i

p3=-2.96983e+002 +7.75369e+002i

p4=-2.96983e+002 -7.75369e+002i

p5=-3.41272e+002 +0.00000e+000i