



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastucture



REF TEK 147A-01/3/EXT P/N:97499-00
S/N: 14701

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0020	1.0020	1.0015
10.00	1.0000	1.0000	1.0000
20.00	0.9944	0.9942	0.9957
30.00	0.9863	0.9859	0.9895
40.00	0.9772	0.9766	0.9827
50.00	0.9689	0.9682	0.9771
60.00	0.9627	0.9620	0.9737
70.00	0.9605	0.9599	0.9742
80.00	0.9635	0.9634	0.9802
90.00	0.9724	0.9731	0.9919
100.00	0.9880	0.9899	1.0103
120.00	1.0399	1.0470	1.0684
140.00	1.1022	1.1189	1.1360
160.00	1.1013	1.1244	1.1342

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
OutputSensitivity (V/g)	2.502	2.501	2.499
Cross-axis Coupling (g/g)	0.48%	0.20%	0.44%