



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
 1600 Tenth Street
 Suite A
 Plano, Texas 75074
 Support phone: 1-888-879-2207
 Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/EXT P/N:97499-00
S/N: 14703

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0031	1.0030	1.0032
10.00	1.0000	1.0000	1.0000
20.00	0.9917	0.9919	0.9914
30.00	0.9810	0.9813	0.9798
40.00	0.9705	0.9713	0.9680
50.00	0.9632	0.9644	0.9589
60.00	0.9602	0.9619	0.9541
70.00	0.9629	0.9649	0.9551
80.00	0.9724	0.9746	0.9629
90.00	0.9880	0.9901	0.9771
100.00	1.0103	1.0120	0.9979
120.00	1.0744	1.0719	1.0577
140.00	1.1421	1.1227	1.1176
160.00	1.1316	1.0810	1.1018

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.500	2.503
Cross-axis Coupling (g/g)	0.24%	0.24%	0.24%