The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrustructure

REF TEK 147A-01/3/EXT P/N:97499-00
S/N: 14714

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0033	1.0032	1.0032
10.00	1.0000	1.0000	1.0000
20.00	0.9912	0.9915	0.9913
30.00	0.9799	0.9802	0.9797
40.00	0.9692	0.9689	0.9676
50.00	0.9618	0.9604	0.9579
60.00	0.9589	0.9558	0.9521
70.00	0.9616	0.9565	0.9514
80.00	0.9712	0.9634	0.9569
90.00	0.9868	0.9761	0.9684
100.00	1.0091	0.9948	0.9860
120.00	1.0731	1.0492	1.0389
140.00	1.1378	1.1048	1.0967
160.00	1.1160	1.0915	1.0925

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
OutputSensitivity (V/g)	2.496	2.502	2.500
Cross-axis Coupling (g/g)	0.76%	0.16%	0.20%