The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure

REF TEK 147A-01/3/EXT P/N:97499-00
S/N: 14720

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0021	1.0017	1.0018
10.00	1.0000	1.0000	1.0000
20.00	0.9940	0.9952	0.9949
30.00	0.9853	0.9881	0.9875
40.00	0.9756	0.9800	0.9791
50.00	0.9670	0.9726	0.9715
60.00	0.9607	0.9669	0.9658
70.00	0.9589	0.9645	0.9637
80.00	0.9627	0.9667	0.9663
90.00	0.9729	0.9743	0.9744
100.00	0.9901	0.9881	0.9885
120.00	1.0461	1.0368	1.0354
140.00	1.1081	1.1040	1.0917
160.00	1.0931	1.1264	1.0935

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
OutputSensitivity (V/g)	2.500	2.500	2.497
Cross-axis Coupling (g/g)	0.16%	0.04%	0.04%