



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
 1600 Tenth Street
 Suite A
 Plano, Texas 75074
 Support phone: 1-888-879-2207
 Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/EXT P/N:97499-00
S/N: 14731

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0019	1.0019	1.0016
10.00	1.0000	1.0000	1.0000
20.00	0.9946	0.9947	0.9956
30.00	0.9869	0.9871	0.9893
40.00	0.9784	0.9786	0.9825
50.00	0.9711	0.9713	0.9769
60.00	0.9665	0.9664	0.9740
70.00	0.9664	0.9657	0.9752
80.00	0.9723	0.9707	0.9823
90.00	0.9846	0.9820	0.9956
100.00	1.0044	1.0003	1.0162
120.00	1.0665	1.0588	1.0804
140.00	1.1355	1.1240	1.1541
160.00	1.1236	1.1106	1.1483

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.503	2.498	2.500
Cross-axis Coupling (g/g)	0.12%	0.20%	0.36%