



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/EXT P/N:97499-00
S/N: 14745

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0019	1.0021	1.0017
10.00	1.0000	1.0000	1.0000
20.00	0.9946	0.9941	0.9953
30.00	0.9867	0.9856	0.9884
40.00	0.9780	0.9759	0.9806
50.00	0.9702	0.9670	0.9734
60.00	0.9648	0.9601	0.9679
70.00	0.9636	0.9573	0.9659
80.00	0.9682	0.9597	0.9685
90.00	0.9793	0.9681	0.9764
100.00	0.9978	0.9834	0.9905
120.00	1.0589	1.0357	1.0391
140.00	1.1292	1.0985	1.1018
160.00	1.1129	1.0933	1.1125

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.500	2.500
Cross-axis Coupling (g/g)	0.28%	0.20%	0.24%