



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
 1600 Tenth Street
 Suite A
 Plano, Texas 75074
 Support phone: 1-888-879-2207
 Support Email: reftek_support@trimble.com
www.trimble.com/infrastucture



REF TEK 147A-01/3/EXT P/N:97499-00
S/N: 14753

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0017	1.0015	1.0017
10.00	1.0000	1.0000	1.0000
20.00	0.9952	0.9957	0.9952
30.00	0.9882	0.9894	0.9882
40.00	0.9805	0.9823	0.9804
50.00	0.9738	0.9759	0.9735
60.00	0.9693	0.9714	0.9686
70.00	0.9688	0.9701	0.9675
80.00	0.9739	0.9734	0.9714
90.00	0.9850	0.9818	0.9811
100.00	1.0032	0.9961	0.9972
120.00	1.0622	1.0432	1.0492
140.00	1.1300	1.0991	1.1060
160.00	1.1166	1.0987	1.0886

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.500	2.503
Cross-axis Coupling (g/g)	0.24%	0.48%	0.60%