The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure

REF TEK 147A-01/3/EXT P/N:97499-00
S/N: 14758

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0016	1.0018	1.0016
10.00	1.0000	1.0000	1.0000
20.00	0.9956	0.9949	0.9954
30.00	0.9892	0.9875	0.9887
40.00	0.9820	0.9792	0.9813
50.00	0.9756	0.9718	0.9748
60.00	0.9711	0.9667	0.9705
70.00	0.9702	0.9655	0.9699
80.00	0.9742	0.9696	0.9747
90.00	0.9838	0.9798	0.9854
100.00	0.9999	0.9970	1.0030
120.00	1.0539	1.0538	1.0608
140.00	1.1231	1.1232	1.1297
160.00	1.1328	1.1239	1.1232

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
OutputSensitivity (V/g)	2.503	2.498	2.503
Cross-axis Coupling (g/g)	0.40%	0.48%	0.28%