



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/EXT P/N:97499-00
S/N: 14768

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0020	1.0020	1.0016
10.00	1.0000	1.0000	1.0000
20.00	0.9944	0.9944	0.9956
30.00	0.9862	0.9864	0.9893
40.00	0.9768	0.9773	0.9823
50.00	0.9680	0.9692	0.9763
60.00	0.9612	0.9633	0.9724
70.00	0.9580	0.9615	0.9722
80.00	0.9596	0.9651	0.9769
90.00	0.9667	0.9748	0.9872
100.00	0.9800	0.9914	1.0036
120.00	1.0263	1.0468	1.0557
140.00	1.0853	1.1130	1.1159
160.00	1.0947	1.1099	1.1147

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
OutputSensitivity (V/g)	2.499	2.499	2.498
Cross-axis Coupling (g/g)	0.56%	0.28%	0.44%