The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure

REF TEK 147A-01/3/EXT P/N:97499-00
S/N: 14771

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0020	1.0016	1.0013
10.00	1.0000	1.0000	1.0000
20.00	0.9944	0.9954	0.9962
30.00	0.9864	0.9887	0.9908
40.00	0.9773	0.9811	0.9848
50.00	0.9692	0.9741	0.9798
60.00	0.9635	0.9690	0.9769
70.00	0.9619	0.9672	0.9776
80.00	0.9662	0.9700	0.9834
90.00	0.9770	0.9781	0.9946
100.00	0.9954	0.9923	1.0124
120.00	1.0575	1.0406	1.0687
140.00	1.1348	1.1024	1.1335
160.00	1.1336	1.1125	1.1240

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
OutputSensitivity (V/g)	2.500	2.499	2.498
Cross-axis Coupling (g/g)	0.32%	0.20%	0.68%