



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
 1600 Tenth Street
 Suite A
 Plano, Texas 75074
 Support phone: 1-888-879-2207
 Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/EXT P/N:97499-00
S/N: 14776

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis	Y-Axis	X-Axis
	(Vertical)	(North/South)	(East/West)
1.00	1.0016	1.0017	1.0013
10.00	1.0000	1.0000	1.0000
20.00	0.9956	0.9952	0.9962
30.00	0.9894	0.9882	0.9908
40.00	0.9825	0.9803	0.9848
50.00	0.9768	0.9733	0.9796
60.00	0.9733	0.9683	0.9763
70.00	0.9739	0.9669	0.9764
80.00	0.9799	0.9705	0.9811
90.00	0.9920	0.9794	0.9910
100.00	1.0111	0.9944	1.0069
120.00	1.0716	1.0416	1.0579
140.00	1.1395	1.0891	1.1193
160.00	1.1251	1.0650	1.1223

Sensitivities

	Z-Axis (Vertical)	X-Axis (East/West)	Y-Axis (North/South)
Output Sensitivity (V/g)	2.501	2.499	2.497
Cross-axis Coupling (g/g)	0.76%	0.20%	0.20%