



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/EXT P/N:97499-00
S/N: 14777

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0018	1.0016	1.0013
10.00	1.0000	1.0000	1.0000
20.00	0.9950	0.9956	0.9963
30.00	0.9879	0.9893	0.9911
40.00	0.9801	0.9824	0.9854
50.00	0.9734	0.9767	0.9808
60.00	0.9693	0.9732	0.9784
70.00	0.9694	0.9738	0.9796
80.00	0.9754	0.9799	0.9860
90.00	0.9878	0.9920	0.9978
100.00	1.0074	1.0110	1.0160
120.00	1.0685	1.0698	1.0710
140.00	1.1300	1.1302	1.1264
160.00	1.0956	1.1008	1.0983

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
OutputSensitivity (V/g)	2.502	2.500	2.499
Cross-axis Coupling (g/g)	0.12%	0.20%	0.36%