The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure

REF TEK 147A-01/3/EXT P/N:97499-00
S/N: 14779

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0030	1.0028	1.0029
10.00	1.0000	1.0000	1.0000
20.00	0.9920	0.9925	0.9921
30.00	0.9815	0.9828	0.9815
40.00	0.9715	0.9736	0.9708
50.00	0.9649	0.9674	0.9626
60.00	0.9634	0.9656	0.9582
70.00	0.9679	0.9692	0.9591
80.00	0.9798	0.9792	0.9662
90.00	0.9977	0.9950	0.9789
100.00	1.0215	1.0169	0.9975
120.00	1.0820	1.0769	1.0508
140.00	1.1305	1.1333	1.1054
160.00	1.1018	1.1111	1.0974

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
OutputSensitivity (V/g)	2.503	2.502	2.499
Cross-axis Coupling (g/g)	0.24%	0.12%	0.08%