The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure

REF TEK 147A-01/3/EXT P/N:97499-00
S/N: 14781

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0017	1.0018	1.0016
10.00	1.0000	1.0000	1.0000
20.00	0.9951	0.9949	0.9956
30.00	0.9881	0.9874	0.9893
40.00	0.9801	0.9789	0.9822
50.00	0.9728	0.9713	0.9758
60.00	0.9674	0.9656	0.9714
70.00	0.9656	0.9637	0.9705
80.00	0.9684	0.9670	0.9742
90.00	0.9767	0.9761	0.9833
100.00	0.9911	0.9921	0.9985
120.00	1.0390	1.0463	1.0480
140.00	1.0957	1.1130	1.1065
160.00	1.0940	1.1107	1.1053

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
OutputSensitivity (V/g)	2.506	2.498	2.504
Cross-axis Coupling (g/g)	0.08%	0.36%	0.44%