



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/EXT P/N:97499-00
S/N: 14800

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0016	1.0018	1.0016
10.00	1.0000	1.0000	1.0000
20.00	0.9954	0.9948	0.9954
30.00	0.9887	0.9873	0.9888
40.00	0.9810	0.9786	0.9814
50.00	0.9741	0.9705	0.9750
60.00	0.9691	0.9641	0.9707
70.00	0.9675	0.9611	0.9701
80.00	0.9705	0.9626	0.9746
90.00	0.9789	0.9693	0.9849
100.00	0.9933	0.9818	1.0016
120.00	1.0417	1.0253	1.0552
140.00	1.1019	1.0788	1.1155
160.00	1.1088	1.0819	1.1046

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.499	2.500	2.501
Cross-axis Coupling (g/g)	0.24%	0.36%	0.40%