The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure

REF TEK 147A-01/3/EXT P/N:97499-00
S/N: 14813

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0017	1.0017	1.0021
10.00	1.0000	1.0000	1.0000
20.00	0.9951	0.9951	0.9942
30.00	0.9882	0.9880	0.9858
40.00	0.9804	0.9800	0.9764
50.00	0.9735	0.9728	0.9679
60.00	0.9689	0.9676	0.9618
70.00	0.9680	0.9660	0.9599
80.00	0.9725	0.9690	0.9635
90.00	0.9827	0.9773	0.9734
100.00	0.9996	0.9914	0.9905
120.00	1.0553	1.0366	1.0479
140.00	1.1247	1.0867	1.1199
160.00	1.1345	1.0808	1.1275

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
OutputSensitivity (V/g)	2.500	2.499	2.498
Cross-axis Coupling (g/g)	0.08%	0.76%	0.40%