



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
 1600 Tenth Street
 Suite A
 Plano, Texas 75074
 Support phone: 1-888-879-2207
 Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/EXT P/N:97499-00
S/N: 14823

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0016	1.0018	1.0015
10.00	1.0000	1.0000	1.0000
20.00	0.9955	0.9948	0.9959
30.00	0.9891	0.9873	0.9900
40.00	0.9818	0.9787	0.9835
50.00	0.9752	0.9709	0.9779
60.00	0.9704	0.9651	0.9744
70.00	0.9690	0.9630	0.9744
80.00	0.9722	0.9660	0.9795
90.00	0.9806	0.9746	0.9898
100.00	0.9948	0.9897	1.0064
120.00	1.0417	1.0402	1.0583
140.00	1.0960	1.1003	1.1145
160.00	1.0907	1.0969	1.0993

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.498	2.500	2.502
Cross-axis Coupling (g/g)	0.48%	0.08%	0.36%