



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/EXT

S/N: 14845

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0021	1.0020	1.0020
10.00	1.0000	1.0000	1.0000
20.00	0.9942	0.9945	0.9944
30.00	0.9859	0.9865	0.9863
40.00	0.9766	0.9777	0.9773
50.00	0.9684	0.9700	0.9694
60.00	0.9627	0.9649	0.9640
70.00	0.9616	0.9642	0.9631
80.00	0.9666	0.9697	0.9684
90.00	0.9787	0.9821	0.9806
100.00	0.9989	1.0027	1.0012
120.00	1.0676	1.0709	1.0706
140.00	1.1553	1.1536	1.1589
160.00	1.1575	1.1439	1.1583

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.500	2.502
Cross-axis Coupling (g/g)	0.16%	0.60%	0.40%